DECLARATION OF CONFORMITY

PVS SpA located in Leonardo da Vinci, 18 street , 20051 Cassina De' Pecchi (Mi), Italy, as assembler of the following kit :

Description	TRIX KIT
REF	CPS011
Basic- UDI-DI	80340280114D – SRN- IT- PR-000050312

Composition :

2 sterile gauze pads cm 18x40 MD class Is

1 pair of scissors cm 10 MD class I

1 plastic tweezers MD class I

1 sterile bandage tnt cm 40x60 MD class Is

1 VENTO mouth to mouth airway MD class IIa

1 pack of 10 adhesive plasters MD class I

1 tissue pack 10 pcs

3 disinfectant wipes;

1 elastic bandage 4x10 cm MD class I

1 adhesive dressing cm 10x6 MD class I

1 adhesive plaster roll m 5 x 1,25 cm MD class I

1 pair of sterile gloves. MD class Is

In compliance with Article 22 of REG. UE 2017/745, declares that :

a) the mutual compatibility of the devices has been verified;

b) is responsible for assembly, packaging and has provided users with the relevant information containing the relevant manufacturer's instructions

(c) the activity of combining the devices into procedural kits has been subjected to adequate methods of internal control, verification and validation.

d) the system incorporates both CE medical devices and non-medical devices.

Cassina De Pecchi, 30/09/25

LEGAL REPRESENTATIVE

IRENE PEREGO