MEDINET srl

SCHEDA TECNICA		
Prodotto	**RESPIPROGRAM**	Data: **01.03.18** Rev.: **06**

Foto	
Destinazione d'uso	Dispositivo per esercitare la respirazione mediante l'inspirazione
Materiale	Base ed indicatori mobili in polipropilene Corpo a tre camere in Polististorolo Tubo e boccaglio in polietilene Tutti i materiali impiegati sono atossici e idonei al contatto con la cavità orale. (Il dispositivo è da intendersi monopaziente).
Imballo	Rif. 15 10 001 001 confezionato in un astuccio Rif. 15 10 002 001 confezionato in sacchetto tipo minigrip Master carton da 24 astucci
Marcatura CE	Marcato Ce classe 1 non sterile senza funzione di misura
Peso	128 g
Stoccaggio	Per garantire l'integrità del prodotto conservare in luoghi asciutti e puliti, lontano da fonti di calore diretto e dalla luce solare.
Avvertenze	Utilizzare questo dispositivo solo sotto la prescrizione e le indicazioni fornite dal medico o l'infermiere. In uso pediatrico, il bambino deve essere sorvegliato durante l'esercizio. Tenere fuori dalla portata dei bambini.

Pulizia	Dopo l'utilizzo lavare accuratamente il boccaglio solo acqua corrente calda (circa 40° C) e lasciare asciugare. É possibile posizionare il tubo con il boccaglio nella sua sede alla destra del corpo, il boccaglio rivolto verso il basso. Respiprogram può essere smontato aprendo la base bianca per il lavaggio delle parti interne.
Smaltimento	Contenitore della plastica.
Note	Latex free
Cnd	Y030327
N. Repertorio	116918
Produttore	Medinet Srl – Milano (Italia)

MEDINET srl

TECHNICAL Data Sheet		
Product	**RESPIPROGRAM** **CODE 1510002001**	Date : **01/03/18** Rev.: **06**

Photo	
Destination of Use	Single patient device to exercise respirations by means of deep inspiration
Materials	Base & moving indicators in polypropylene Three chambers body in Polystyrene Tube & mouthpiece of polyethylene All materials used are no-toxics and suitable for contact with the oral cavity (the device is intended single patient).
Packaging	Packaged in a minigrip plastic bag Master carton from 24 bags
CE CLASS	Class I non sterile without measuring function
Weight	128 g
Storage	To ensure the integrity of the product store in a clean and dry place, away from direct heat and from sunlight.
Warnings	This device has to be used only under the prescription and the indications given by the doctor or nurse. For pediatric use, the child should be monitored during the exercise. Keep out of the reach of children.

This document is the exclusive property of Medinet Srl Milan. Reproduction / unauthorized distribution even partial, will be prosecuted according to law

Cleaning	After use wash carefully the mouthpiece only hot running water (about 40° C) and leave to dry. It's possible to place the tube with the mouthpiece in its seat to the right of the body, the mouthpiece, facedown. Respiprogram can be removed opening the white base for the washing of the internal parts.
Disposing	Plastic cycle.
Notes	Latex free
Cnd	Y030327
Repertory No.	116918
Supplier	Medinet Srl – Milano (Italy)

This document is the exclusive property of Medinet Srl Milan. Reproduction / unauthorized distribution even partial, will be prosecuted according to law