**Quincy Lab, Inc.**
109 Shore Drive
Burr Ridge, IL 60527

CE 2024

# EC DECLARATION OF CONFORMITY

We, QUINCY LAB, INC.
109 Shore Drive
Burr Ridge, IL 60527 USA

Contact: Cheri McKown
Phone: 800-482-4328
Fax: 773-622-2282
E-mail: cheri@quincylab.com

declare under our sole responsibility, that our Incubator/Warmer, models as listed below: (*but not limited to*)

| | | | | | | |
|---|---|---|---|---|---|---|
| **10-140-1** | **12-140-1** | **12-140AEr-1** | **10-180E-1** | **12-180-1** | **12-180AEr-1** | **12-140AEC-1** |
| **10-140E-1** | **12-140E-1** | **12-140AEg-1** | **10-180AE-1** | **12-180E-1** | **10-140AEs-1** | **10-140MHSe-1** |
| **10-140AE-1** | **12-140AE-1** | **10-180-1** | **10-180AEr-1** | **12-180AE-1** | **12-140AEs-1** | **12-140MHSe-1** |
| **10-140MHSv-1** | **10-140MHSm-1** | **12-180AEm-1** | | | | |

to which this declaration relates, are in conformity with the standards listed below:

- **IEC 61010-1:2010, AMD1:2016** – Safety requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General requirements
- **IEC 61010-2-010:2014** – Safety requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-010: Particular Requirements for Laboratory Equipment for the Heating of Material (for use in conjunction with **IEC 61010 1:2010**)

**Verification Issuing Office:** Intertek (NRTL)

**Verification number(s):** 103877412CHI-001VoC, 103877412CHI-002-VoC

**Per Report number(s):** 103877412CHI-001, 103877412CHI-002



Cheri McKown, President

Date: 9/18/2024

**QUINCY LAB, INC.**
1925 N LEAMINGTON AVENUE
CHICAGO, IL 60639

To: **ATTN: Terezia**

From: Ron Modica

Date: 1/27/21

Re: **Quincy / Medline Model No. Cross Reference**

---


Hello Terezia,

Below is the cross-reference verification for the Medline Model Warmers.

| Medline Model No | Quincy Lab Model No |
|---|---|
| MSCEWARMER14S | 10-140MHSe-1 |
| MSCEWARMER28S | 12-140MHSe-1 |

Please let me know if you have any further needs.

Regards,



***Ron Modica***
Project Engineer

**Quincy Lab, Inc.**
1925 N. Leamington Av | Chicago, IL 60639
P 773.622.2428| F 773.622.2282
ron@quincylab.com
***Visit us at*** www.quincylab.com